

AED Inspection Report 10-30-25

Last Inspected	Facility	Address	Model#	Serial#	Adult pad exp.1	Adult pad exp.2	Pedi exp	Battery capacity%	Battery replacement date	Response Ready	Actions
10-30-25	Clubhouse WM	199 Sloan Road, Phoenixville, PA 19460	AED Plus	X23C588909	11-06-27	na	na	60	05-11-28	Yes	delete
10-30-25	Right Stairwell Lower Level	608 Regency Hill Drive, Collegeville, PA 19426	AED Plus	X23B587417	10-28-27	na	na	60	05-11-28	Yes	delete

